

Letter of Declaration

To whom it may concern,

We, Hoyotek Biomedical Co., Ltd., located at No. 225 Jinger Road, Tianjin Airport Economic Zone, China, hereby declare that our Product of Corona Virus (COVID-19) Antigen Rapid Test (Colloidal Gold) have a limit of detection for COVID-19 Antigen was 200 TCID50/ml.

Experiments were conducted on product of the Corona Virus (COVID-19) Antigen Rapid Test (Colloidal Gold) using inactivated SARS-CoV-2 Virus. The experiment result proves that the product have a good sensitivity and meet the required standards.

Yours sincerely,

天津鸿宇泰生物科技有限公司
Hoyotek Biomedical Co., Ltd.
Hoyotek Biomedical Co., Ltd.

Date: December 14, 2021